



**Scientific
Systems
Design Inc.**

Innovative Engineering for Science

Digitimer Limited
37 Hydeway
Welwyn Garden City
AL7 3BE, UK
Tel: +44 (0)1707 328347
Email: sales@digitimer.com

Microscope Stage Slice Chamber

MS4


Submerged Preparations - Inverted Microscope Chamber

The MS-4 chamber has been designed for use with all inverted microscope stages. The simple design consists of an input and output channel arranged so that the glass coverslip area remains filled with perfusion solution. The slices rest on the glass coverslip arranged to be at the focal plane of microscope objective and access for condenser optics from above. Adaptor plates are available for various microscope stages.

FEATURES

- * 1 ml capacity flow through upright microscope stage chamber
- * Glass coverslip base
- * Temperature control unit available

Perfusion solution enters from one side via a stainless steel tube and exits through an 'L' shaped well in which a movable tube attached to a pump / suction line adjusts the fluid level. Solutions can be pre-warmed before entering the chamber with our Mini Heating Module **MH02**. Perfusion solution forms a pool around the preparation whilst being refreshed by virtue of the inlet and exit ports.



SPECIFICATIONS

Size: 100mm diameter X 10mm

Material: Clear acrylic

Coverslip base: 22mm X 150um .

Inlet tube: Stainless steel tube, 1.8 mm O.D.

Exit tube: As for inlet tube

Dead space: Min. capacity of pool formed above coverslip and channels leading to inlet and exit ports is 1ml.



**Scientific
Systems
Design Inc.**

Innovative Engineering for Science

Digitimer Limited
37 Hydeway
Welwyn Garden City
AL7 3BE, UK
Tel: +44 (0)1707 328347
Email: sales@digitimer.com

Microscope Stage Slice Chamber

MS4

Submerged Preparations - Inverted Microscope Chamber

FREQUENTLY ASKED QUESTIONS

My perfusion solution forms large globules around the center of the chamber and has difficulty forming a continuous inflow-outflow stream, why does this happen?

This will happen with a new chamber or one that has just been thoroughly cleaned. The polished surfaces of acrylic are hydrophobic, so solutions tend to form globules rather than forming a meniscus with the walls and channels of the chamber. To overcome this problem for a new chamber or one that has been thoroughly cleaned, leave a saline (not ACSF) solution to form a pool in the chamber overnight. Rinse as normal before use for the next experiment. We do not have an explanation for this phenomenon but know that it does work for most situations. An alternative approach is to cut strips of lens tissue 5mm in width and lay these along the inlet and exit channels to encourage flow across. In some situations it may be necessary to place a piece of lens tissue at the mouth of the exit tube to encourage smooth outflow if the peristaltic pump is too pulsatile.

How do you hold slices down?

The most reliable method is to prepare a U-shaped piece of stainless steel wire, which is then partially flattened to form a suitable surface for adhesive. Thin strands of nylon such as those obtained by separating out the individual fibres in a length of string from dental floss are positioned taut across the U-shape. Cyanoacrylate adhesive is then applied to the flattened surface on which the fine nylon threads are held taut. Once completely dry the excess ends of the fibres are cut. This "harp" is then sunk over the submerged slice preparation. We are hoping to manufacture these at some point.

What is the best method of cleaning the chamber after use?

Rinse with plenty of distilled water and leave dry before the next experiment. NEVER use a solvent of any kind as acrylics tend to fragment and can be completely destroyed. A mild acid (citric) should be used to get rid of deposits, and then left in a large volume of distilled water overnight.

How do I replace the coverslip if it should break?

Wear safety goggles for this procedure. First place the chamber upside down on a flat, soft surface (cork tile) and GENTLY remove all the broken glass, using a small flat ended spatula in a scraping action to remove the old silicone rubber sealant. DO NOT use any kind of solvent to remove the sealant. Once completely clear, use a syringe loaded with silicone sealant and bead this around the circumference of the indentation. Place a clean 22mm coverslip on the fresh sealant and GENTLY press around the circumference ONLY of the coverslip. It is preferable to use a black sealant as it allows the thickness of the sealant to be controlled: darker areas have too much sealant under the coverslip. Try to ensure the final position of the coverslip is central and also very slightly BELOW the UNDERSIDE surface of the chamber, so that when it rests on the microscope stage, the coverslip does not rub on the stage surface. The indentation for the coverslip is 300um in depth, the coverslip supplied with the chamber is 150um, giving room for a good 100um thickness of sealant. Remove excess sealant whilst still wet, taking care not to displace the coverslip. Alternatively return the chamber to us and we will replace the coverslip, please contact us for the cost of this service and further details.

Where is the best electrical ground point?

This will have to be found by trial and error. If you use the original stainless steel inlet and outlet tube as a ground, it will also help to eliminate any artefacts from the suction line. Quite often it helps to push a grounding wire INTO the wall of silicone rubber tubes that you use for connection to your perfusion solutions.